

# EU-KONFORMITÄTSERKLÄRUNG

## EU DECLARATION OF CONFORMITY

Name und Anschrift des Importeurs und Händlers in Europa  
*Name and address of importer and distributor responsible in Europe*

EUE BV  
Derde Oosterparkstraat 73-1  
1091 JV Amsterdam  
Netherlands

DIESE KONFORMITÄTSERKLÄRUNG WIRD UNTER DER ALLEINIGEN VERANTWORTUNG ERSTELLT VON:  
*THIS DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE RESPONSIBILITY OF:*

Name und Anschrift des Herstellers:  
*name and address of manufacturer:*

Changzhou Jinpeng Medical Equipment Co., Ltd  
Adresse: Yanling Village, Zhenglu Town, Tianning District,  
Changzhou, Jiangsu, China

Produktidentifikation:  
*Product identification:*

Markenname/ **HUYOUCZ**  
*brandname*  
Modell. Nr./ **KZ-JP-95-B**  
*model.no.*

Die Konformitätsbewertung nach den wesentlichen Anforderungen wurde erstellt von Jinpeng gemäß  
*The conformity assessment according to the essential requirements has been done by Jinpeng according to*

EU-Rechtsvorschriften der Gemeinschaft:  
*EU-Community Legislation:*

Verordnung über persönliche Schutzausrüstung (PSA) (EU) 2016/425  
*Personal Protective Equipment (PPE) Regulation (EU) 2016/425*  
Anhang II Kategorie III in Verbindung mit  
*Annex II Category III in connection with*  
Anhang V (Modul B EU-Baumusterprüfungsbescheinigung)  
*Annex V (Module B EU Type Examination Certification)*  
A18/000111 und  
Anhang VII (Modul D Qualitätssicherung Produktion)  
*Annex VII (Module D Quality assurance system of the production process)*  
A18/000197

Harmonisierte Normen:  
*harmonized standards:*

Persönliche Schutzausrüstung  
**EN 149:2001+A1:2009**  
*Personal protective equipment*  
**EN 149:2001+A1:2009**  
Jiangsu Guojian Testing Technology Co., Ltd  
3/F., Unit D, Xingye Building, Taihu International Tech-Park, Wuxi, Jiangsu, China  
Report No.:(2020) WSZ FHL NO.6737

Weitere Spezifikationen:  
*further specifications:*

Partikelfiltermaske, Kategorie III, FFP2 NR  
*Particle filtering half mask, Category III, FFP2 NR*

In Zusammenarbeit mit Benannter Stelle:  
*in corporation with:*

AENOR INTERNACIONAL S.A.U (CE 0099)  
Genova, 6 28004 Madrid Espana  
*Notified Body:*

Diese Erklärung ist gültig ab 19. 08. 2020 bis zum 19. 08. 2025  
*This declaration is valid from 19.08.2020 until 19.08. 2025*

UNTERZEICHNET FÜR UND IM NAMEN VON:  
*SIGNED FOR AND BEHALF OF:*

Ort und Datum der Ausstellung  
*Place and date of issue:*

19.08.2020 Changzhou

Unterschrift  
*Signature:*

Name, Funktion  
*Name and function:*

Geschäftsführer  
Jiannan Yao

Firmenname/Company name:

Changzhou Jinpeng Medical Equipment Co., Ltd



# AENOR

## Certificado de Examen UE de Tipo EU Type-Examination Certificate

**A18/000111**

AENOR, como organismo notificado (n° 0099) para el Reglamento (UE) 2016/425, ha emitido este certificado a favor de  
In compliance with Regulation (EU) 2016/425, the notified body AENOR (n° 0099) has issued this certificate to

### **Changzhou Jinpeng Medical Equipment Co., Ltd**

Domicilio social / Registered office No. 715 Heheng Road, Yanling Village, 213115 Zhenglu Town, Tianning District, Changzhou City, Jiangsu Province, (China)

para el producto / for the product Dispositivos de protección respiratoria. Medias máscaras filtrantes de protección contra partículas. / Respiratory protection devices. Half filter masks to protect against particles.

conforme con el Reglamento in compliance with Regulation Reglamento UE 2016/425 de Equipos de Protección Individual (Regulation EU 2016/425 on Personal Protective Equipment)

Norma armonizada / Harmonized standard EN 149:2001+A1:2009

Más información en el anexo / See annex for more information.

Centro de producción / Production site No. 715 Heheng Road, Yanling Village, 213115 Zhenglu Town, Tianning District, Changzhou City, Jiangsu Province, (China)

Esquema de evaluación Assessment scheme Anexo V (Examen UE de Tipo – Módulo B) del Reglamento (UE) 2016/425.

Annex V (EU Type-examination – Module B) of Regulation (EU) 2016/425.

Fecha de emisión / First issued on 2020-08-19  
Fecha de expiración / Validity date 2025-08-19



Rafael GARCÍA MEIRO  
Director General / CEO

Original Electronic Certificate

# AENOR

## Certificado de Examen UE de Tipo EU Type-Examination Certificate

A18/000111

### Anexo al Certificado Annex to Certificate

Norma armonizada / Harmonized standard EN 149:2001+A1:2009

Marca Comercial / Trade Mark	Referencia / Reference	Clasificación / Classification	Descripción / Description
HUYOUCZ	KZ-JP-95-B	FFP2 NR	MEDIA MASCARILLA, DE CINCO CAPAS DE FILTRADO, DOS LAZOS FIJOS DE SUJECCIÓN A OREJAS, DE TIPO PLEGABLE. DISEÑADA PARA PROTEGER CONTRA PARTÍCULAS SÓLIDAS O LÍQUIDAS SUSPENDIDAS EN EL AIRE. NO REUTILIZABLE / FILTERING HALF MASK, FIVE FILTERING LAYERS, TWO EARLOOPS, FOLDING STYLE. DESIGNED TO PROTECT AGAINST AIRBORNE SOLID OR LIQUID PARTICLES. NON-REUSABLE

Fecha de emisión / First issued on 2020-08-19  
Fecha de expiración / Validity date 2025-08-19

Original Electronic Certificate

AENOR INTERNACIONAL S.A.U.  
Génova, 6. 28004 Madrid. España  
Tel. 91 432 60 00.- www.aenor.com

Organismo de control acreditado por ENAC con acreditación N° 1/C-PR354  
Control body accredited by ENAC. Accreditation number 1/C-PR354

# AENOR

## Certificado de Conformidad Certificate of Conformity


CE  
0099

**A18/000197**

AENOR, como organismo notificado (nº 0099) para el Reglamento (UE) 2016/425, ha emitido este certificado a favor de  
In compliance with Regulation (EU) 2016/425, the notified body AENOR (nº 0099) has issued this certificate to

### Changzhou Jinpeng Medical Equipment Co., Ltd

Domicilio social / Registered office	No. 715 Heheng Road, Yanling Village, 213115 Zhenglu Town, Tianning District, Changzhou City, Jiangsu Province, (China)
para aprobar el in order to approve the	Sistema de aseguramiento de la calidad del proceso de producción (módulo D) Quality assurance system of the production process (module D)
conforme con el in compliance with	Reglamento (UE) 2016/425, Anexo VIII Regulation (EU) 2016/425, Annex VIII
Referencias / References	Detalladas en el Anexo al Certificado / Specified in Annex to the Certificate
Centro de producción / Production site	No. 715 Heheng Road, Yanling Village, 213115 Zhenglu Town, Tianning District, Changzhou City, Jiangsu Province, (China)
Esquema de evaluación Assessment scheme	Este certificado se limita al sistema de aseguramiento de la calidad del proceso de producción para los equipos amparados por los certificados de examen UE de tipo detallados en el anexo a este Certificado y fabricados en el centro indicado más arriba.  This certificate is exclusively limited to the quality assurance of the production process for personal protective equipment covered by the EU type-examination certificates detailed in annex to the present certificate and to the above mentioned production site.
Fecha de emisión / First issued on Fecha de expiración / Validity date	2020-10-05 2023-10-05



Rafael GARCÍA MEIRO  
Director General / CEO

Original Electronic Certificate

AENOR INTERNACIONAL S.A.U.  
Génova, 6. 28004 Madrid. España  
Tel. 91 432 60 00.- www.aenor.com

Organismo de control acreditado por ENAC con acreditación Nº 1/C-PR354  
Control body accredited by ENAC. Accreditation number 1/C-PR354





# CERTIFICATE OF APPRECIATION

This is to certify that the Quality Management System of

**[Changzhou Jinpeng Medical Equipment Co., Ltd]**

Unified Social Credit Code: 91320402745551034J

Operation Address : No.715, Heheng Road, Yanling village, Zhenglu Town, Tianning District, Changzhou City, Jiangsu Province, China

Registered Address : No.715, Heheng Road, Yanling village, Zhenglu Town, Tianning District, Changzhou City, Jiangsu Province, China

applicable to

**[Production and sales of daily masks (non-medical)]**

has been assessed and registered by SIRIM against the provisions of

**ISO 9001:2015**

This registration is subject to the company maintaining a Quality Management System to the above standard, which will be monitored by SIRIM.

The validity of this certificate may be verified on both CNCA's website ([www.cnca.gov.cn](http://www.cnca.gov.cn)) & SIRIM's website: [www.sirim-global.com](http://www.sirim-global.com)

Certificate No: 8120Q10181R0S

Initial Certification: 2020.09.01

Issue/Reissue Date: 2020.09.01

Registration Valid Until: 2023.08.31

Certificate Expiry: 2021.08.31



Certification Check

Signed by 





中国认可  
国际互认  
检测  
TESTING  
CNAS L10118



国检检测  
CHINA COMPONENTS TEST

# Test Report

Report No.: [2020] WSZ FHL NO.6737

Product Name Particle filtering half mask

Applicant Changzhou Jinpeng Medical Equipment Co.,Ltd.

Manufacturer Changzhou Jinpeng Medical Equipment Co.,Ltd.

Test Type Entrusted inspection

Jiangsu Guojian Testing Technology Co., Ltd.  
3/F., Unit D, Xingye Building, Taihu International Tech-Park, Wuxi, Jiangsu, China

检验专用章

# Test Report

Product name	Particle filtering half mask	Model name	KZ-JP-95-B
		Brand	HUYOUCZ
Laboratory/ Add.	Jiangsu Guojian Testing Technology Co., Ltd./ 3/F., Unit D, Xingye Building, Taihu International Tech-Park, Wuxi, Jiangsu, China		
Applicant/ Add/Tel	Changzhou Jinpeng Medical Equipment Co.,Ltd./715 Heheng Road, Yanling Village, Zhenglu Town, Tianning District, Changzhou City, Jiangsu Province, China./13861225085		
Manufacturer/ Add/Tel	Changzhou Jinpeng Medical Equipment Co.,Ltd./715 Heheng Road, Yanling Village, Zhenglu Town, Tianning District, Changzhou City, Jiangsu Province, China./13861225085		
Sample classification	FFP2	Sample number	GW6737-2020
Sample quantity	110 pcs	Date of receipt of sample	17/06/2020
Test type	Entrusted inspection	Article/Batch/Style number	—
Date (s) of performance of tests	17/06/2020~26/06/2020	Testing location	Same as the Laboratory
Sample state	Meeting the requirements of testing	Sample description	Refer to page 3
Test standard(s)	EN 149:2001+A1:2009 Respiratory protective devices - Filtering half masks to protect against particles - Requirements, testing, marking		
Test items	Packaging, material, practical performance, finish of parts, compatibility with skin, flammability, carbon dioxide content of the inhalation air, head harness, field of vision, penetration of filter material, breathing resistance, total inward leakage		
Test conclusion	The samples upon testing comply with FFP2 classification requirements according to the standard EN 149:2001+A1:2009. The details of test results see on Pages 3-11. Date of issue: 26/06/2020		
Note	The test results presented in this report relate only to the submitted sample as received.		

Lu Bing

Approver (name, signature)

Wan Heng

Reviewer (name, signature)

Yang Ying

Chief Tester (name, signature)



<b>Sample description:</b>	—
<b>Test item particulars:</b>	
Type of use .....	<input type="checkbox"/> re-useable particle filtering half mask <input checked="" type="checkbox"/> single shift only particle filtering half mask
Classes of devices.....	<input type="checkbox"/> FFP1 <input checked="" type="checkbox"/> FFP2 <input type="checkbox"/> FFP3
Exhalation valve(s).....	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
Inhalation valve(s).....	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
Designed to protect against both solid & liquid aerosols. :	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
<b>Possible test case verdicts:</b>	
- Test case does not be required to the test object.....: NRq (Not required)	
- Test case does not apply to the test object.....: N/A (Not Applicable)	
- Test object does meet the requirement.....: P (Pass)	
- Test object does not meet the requirement.....: F (Fail)	
<b>General remarks:</b>	
<p>The test results presented in this report relate only to the submitted sample as received.</p> <p>This report shall not be reproduced, except in full, without the written approval of the issuing Laboratory can provide assurance that parts of a report are not taken out of context.</p> <p>Determination of the test results includes consideration of measurement uncertainty from the test equipment and methods.</p> <p>Throughout this report a <input type="checkbox"/> comma / <input checked="" type="checkbox"/> point is used as the decimal separator.</p>	
<b>Environmental condition of the testing in this report:</b>	
1) Unless otherwise specified, the ambient temperature for testing shall be 25 °C;	
2) T.C. Temperature conditioned:	
a) for 24 h to a dry atmosphere of 70 °C;                      b) for 24 h to a temperature of -30 °C;	
and return to room temperature 25 °C for 4 h between exposures and prior to subsequent testing.	

S.No. (Cl.No.)	Test item		Unit	Technical requirements	Test result	Single item decision
1 (7.3)	Visual inspection	Marking/ information	—	Marking and the information supplied by the manufacturer, requirements refer to Cl.9 and Cl.10	The clause were not required	NRq
2 (7.4)	Packaging	Visual inspection	—	Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.	Particle filtering half masks packaged and protected against mechanical damage and contamination.	Pass
3 (7.5)	Material	Visual inspection	—	Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used.	Materials were suitable withstand handling and wear.	Pass
			—	After undergoing S.W., none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps.	Sample 1: neither facepiece nor straps have mechanical failure	
			—		Sample 2: neither facepiece nor straps have mechanical failure	
			—		Sample 3: neither facepiece nor straps have mechanical failure	
			—	After undergoing S.W. and T.C., none of the particle filtering half masks shall not collapse.	Sample 4: no collapse	
—	Sample 5: no collapse					
—	Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer.	Sample 6: no collapse				
—		Not constitute a hazard or nuisance for the wearer				
4 (7.6)	Cleaning and disinfecting		—	Particle filtering half mask designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer. Testing shall be done in accordance with 8.4 and 8.5.	<input type="checkbox"/> Fulfil the requirements after testing, or <input checked="" type="checkbox"/> The Particle filtering half mask is NOT re-usable according to information supplied by manufacturer	N/A
			—	With reference to 7.9.2, after cleaning and disinfecting the re-usable particle filtering half mask shall satisfy the penetration requirement of the relevant class. Testing shall be done in accordance with 8.11.	<input type="checkbox"/> Tests results refer to S. No. 7(7.9.2), or <input checked="" type="checkbox"/> The Particle filtering half mask is NOT re-usable according to information supplied by manufacturer	

S.No. (Cl.No.)	Test item	Unit	Technical requirements	Test result	Single item decision		
5 (7.7)	Practical performance	Head harness comfort	—	Head harness should be comfort.	Sample 1: has the feeling of comfortable wearing	Pass	
					Sample 2: has the feeling of comfortable wearing		
		Security of fastenings	—	Fastenings are safe and reliable	Sample 1: All fastenings are firm		Sample 2: All fastenings are firm
					Sample 2: All fastenings are firm		
		Field of vision	—	Field of vision is acceptable	Sample 1: Having a wider visual field		Sample 2: Having a wider visual field
					Sample 2: Having a wider visual field		
6 (7.8)	Finish of parts	Visual inspection	—	Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.	Parts of the device have no sharp edges and burrs	Pass	
7 (7.9.2)	Leakage— Penetration of filter material	Sodium chloride	—	$\leq 6\%$	A.R. <sup>1)</sup> 0.1% 0.1% 0.1%	Pass	
					S.W. <sup>1)</sup> 0.1% 0.1% 0.1%		
					M.S+ T.C. <sup>2)</sup> 0.2% 0.2% 0.2%		
		Paraffin oil	—	$\leq 6\%$	A.R. <sup>1)</sup> 0.3% 0.4% 0.4%	Pass	
					S.W. <sup>1)</sup> 0.4% 0.4% 0.3%		
					M.S+ T.C. <sup>2)</sup> 1.0% 1.2% 1.1%		
		<sup>1)</sup> average penetration over a time of 30s, beginning 3 min after the start of the test reported <sup>2)</sup> max. penetration during exposure test reported; Note: The penetration of the filter of the particle filtering half mask shall meet the requirements below: Maximum penetration of sodium chloride aerosol test 95 L/min max. FFP1: 20%, FFP2: 6%, FFP3: 1% Maximum penetration of paraffin oil aerosol test 95 L/min max. FFP1: 20%, FFP2: 6%, FFP3: 1%					

S.No. (Cl.No.)	Test item	Unit	Technical requirements	Test result		Single item decision
8 (7.10)	Compatibility with skin	—	Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.	A.R.	5 pcs all don't cause irritation	Pass
				T.C.	5 pcs all don't cause irritation	
9 (7.11)	Flammability	—	When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5s after removal from the flame.	A.R.	The Sample is burning. Burning time:0.1s	Pass
					The Sample is burning. Burning time:0.1s	
				T.C.	The Sample is burning. Burning time:0.1s	
					The Sample is burning. Burning time:0.1s	
10 (7.12)	Carbon dioxide content of the inhalation air	—	The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1.0 % (by volume). Remark: 3 half masks (S1, S2 and S3) A.R. tested.	Sample 1	0.7224%	Pass
				Sample 2	0.7218%	
				Sample 3	0.7230%	
				average	0.72%	
11 (7.13)	Head harness	—	The head harness shall be designed so that the particle filtering half mask can be donned and removed easily. The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining total inward leakage requirements for the device.	A.R.	All of 5 pieces particle filtering half mask meet the requirements	Pass
				T.C.	All of 5 pieces particle filtering half mask meet the requirements	
12 (7.14)	Field of vision	—	The field of vision is acceptable if determined so in practical performance tests.	The two samples both have a wider visual field		Pass

S.No. (Cl.No.)	Test item	Unit	Technical requirements	Test result	Single item decision
13 (7.15)	Exhalation valve(s)	—	A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations.	No exhalation valve(s)	N/A
		—	If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage, and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9.	No exhalation valve(s)	
		—	Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30 s.	No exhalation valve(s)	
		—	When the exhalation valve housing is attached to the faceblank, it shall withstand axially a tensile force of 10 N applied for 10 s.	No exhalation valve(s)	
14 (7.17)	Clogging— Breathing resistance & Penetration of filter material	—	Optional for single shift use devices, mandatory for re-usable devices. Tested by Cl. 7.17.1/2/3.	<input type="checkbox"/> Tests results refer to Table C&D, or <input checked="" type="checkbox"/> Tests not requested for single shift use face mask	N/A
15 (7.18)	Demountable parts	—	All demountable parts (if fitted) shall be readily connected and secured, where possible by hand.	No demountable parts	N/A

**Table A- Leakage—Total Inward Leakage**

S.No. (Cl.No.)	Test item	Unit	Technical requirements <sup>1)</sup>	Test result						Single item decision	
				Exercises	E1 (%)	E2 (%)	E3 (%)	E4 (%)	E5 (%)		TIL (%)
16 (7.9.1)	Leakage— Total inward leakage	—	At least 46 out of the 50 individual exercise results shall be not greater than <b>11%</b> ; And in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than <b>8%</b> .	A.R.	4.5	5.0	5.8	6.0	4.5	5.2	Pass
					5.2	6.6	7.2	7.3	5.7	6.4	
					4.8	5.7	5.8	6.2	4.7	5.4	
					5.4	6.2	6.7	6.9	4.9	6.0	
					4.8	5.7	6.0	6.2	4.3	5.4	
				T.C.	5.2	6.3	7.0	7.2	5.7	6.3	
					4.2	5.5	5.6	5.8	4.6	5.1	
					4.7	5.4	5.6	6.2	4.2	5.2	
					4.8	5.7	5.9	6.1	4.6	5.4	
					5.2	5.8	6.4	6.6	4.7	5.7	

**Note 1:**

at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than 25 % for FFP1 11 % for FFP2 5 % for FFP3

in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than 22 % for FFP1 8 % for FFP2 2 % for FFP3.

**Table A-1- Test subjects—Facial dimension**

Test Subject No.	Length of face (mm)	Width of face (mm)	Depth of face (mm)	Width of mouth (mm)
1	120	130	109	59
2	122	140	115	65
3	119	160	139	55
4	112	122	119	63
5	110	130	118	60
6	115	119	110	59
7	112	123	113	55
8	103	130	100	50
9	118	139	130	63
10	120	135	125	50

**Table B- Breathing Resistance**

S.No (Cl.No.)	Test item		Unit	Technical requirements <sup>1)</sup>	Test result					Single item decision	
					Exercises	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side		Lying on the right side
17 (7.16)	Breathing resistance	Inhalation 30 L/min	mbar	$\leq 0.7$	A.R.	0.6	0.5	0.6	0.5	0.6	Pass
						0.5	0.6	0.6	0.6	0.6	
						0.6	0.6	0.5	0.6	0.5	
					S.W.	0.6	0.6	0.5	0.6	0.6	
						0.6	0.5	0.6	0.6	0.5	
						0.5	0.6	0.6	0.5	0.6	
					T.C.	0.6	0.6	0.5	0.6	0.6	
						0.6	0.5	0.6	0.6	0.5	
						0.5	0.6	0.6	0.5	0.6	
	Breathing resistance	Inhalation 95 L/min	mbar	$\leq 2.4$	A.R.	1.8	1.7	1.8	1.8	1.7	Pass
						1.8	1.8	1.7	1.8	1.8	
						1.7	1.8	1.8	1.7	1.8	
					S.W.	1.8	1.7	1.8	1.8	1.7	
						1.7	1.8	1.8	1.7	1.8	
						1.8	1.8	1.7	1.8	1.8	
					T.C.	1.7	1.8	1.8	1.7	1.7	
						1.8	1.8	1.7	1.8	1.8	
						1.8	1.7	1.8	1.8	1.8	
Breathing resistance	Exhalation 160 L/min	mbar	$\leq 3.0$	A.R.	2.6	2.7	2.6	2.6	2.6	Pass	
					2.6	2.6	2.6	2.6	2.7		
					2.7	2.6	2.7	2.5	2.6		
				S.W.	2.6	2.7	2.6	2.6	2.6		
					2.7	2.6	2.6	2.7	2.6		
					2.6	2.6	2.7	2.6	2.5		
				T.C.	2.6	2.7	2.6	2.6	2.7		
					2.7	2.6	2.6	2.5	2.6		
					2.6	2.6	2.7	2.6	2.6		

Note 1: Limitation may need be changed according to classification, refer to Table 2 — Breathing resistance of EN 149:2001 +A1:2009 for the Technical requirements.

**Table C- Clogging Test—Breathing resistance**

S.No (CLNo.)	Test item <sup>1)2)</sup>		Unit	Technical requirements <sup>1) 2)</sup> (mbar)	Test result						Single item decision
					Exercises	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side	Lying on the right side	
18 (7.17)	Clogging test—	Inhalation 95 L/min	mbar	—	A.R.						N/A
					T.C.						
	Breathing resistance	Exhalation 95 L/min	mbar	—	A.R.						N/A
						T.C.					

Note 1: Valved particle filtering half masks

After clogging the inhalation resistances shall not exceed FFP1: 4 mbar FFP2: 5 mbar FFP3: 7 mbar at 95 L/min continuous flow;  
The exhalation resistance shall not exceed 3 mbar at 160 L/min continuous flow.

Note 2: Valveless particle filtering half masks

After clogging the inhalation and exhalation resistances shall not exceed FFP1: 3 mbar, FFP2: 4 mbar FFP3: 5 mbar at 95 L/min continuous flow.

**Table D- Clogging Test—Penetration of filter material**

S.No (CLNo.)	Test item	Unit	Technical requirements	Test result		Single item decision
19 (7.17)	Clogging test- Penetration of filter material	Paraffin oil	—	—	A.R.	N/A
					T.C.	
					T.C.	

Note: Maximum penetration of test aerosol test 95 L/min max. FFP1: 20%, FFP2: 6%, FFP3: 1%

**Abbreviations :**


A.R. As received	M.S. Mechanical strength	S.W. Simulated wearing treatment
T.C. Temperature conditioned	F.C. Flow conditioned	C.D. Cleaning and Disinfecting



**Annex A- Estimates of the uncertainty of measurement**

Test item	Uncertainty
Total inward leakage	2.98%
Penetration of filter material	1.00%
Flammability	1.00%
Carbon dioxide content of the inhalation air	0.93%
Breathing resistance	1.90%

**Annex B- Sample Photo**



The end




171021110579



中国认可  
国际互认  
检测  
TESTING  
CNAS L7901

# 检验检测报告

## TEST REPORT



STFWT20204824

Product Name	Particle Filtering Half Mask
Trust Unit	Changzhou Jinpeng Medical Equipment Co., Ltd
Manufacturer	Changzhou Jinpeng Medical Equipment Co., Ltd
Test Category	Trust Test




江苏省特种安全防护产品质量监督检验中心  
JIANGSU QUALITY SUPERVISION AND INSPECTION CENTER FOR SPECIAL SAFETY PROTECTION PRODUCTS

检验检测中心

# Test Report

STFWT20204824

Page 1 of 5

Product Name	Particle Filtering Half Mask	Specification Type	Adult
		Trademark	---
Trust Unit	Changzhou Jinpeng Medical Equipment Co., Ltd	Tel	13861225085
Manufacturer	Changzhou Jinpeng Medical Equipment Co., Ltd	Sample Grade	KN95
Sample Quantity	60 只	Sample Receiving Date	2020-03-13
Test Category	Trust Test	Serial Number	20200308
Samples Conditions	Accord with detecting request		
Document and Decide Accordance	GB 2626-2006 《Respiratory Protective Apparatus Self-suction filtered anti-particulate respirator》		
Test Conclusion	The samples were tested, and the items tested meet the requirements of GB 2626-2006 standard KN95 level. <div style="text-align: right;">  <p>Signature Date: 2020-03-24</p> </div>		
Remarks	This report is responsible for samples only.		

Approver

*陈敏*

Examiner

*杨森*

Majortester

*熊永涛*

## Testing Results

STFWT20204824

Page 2 of 5

Serial	Test Items	Unit	Requirement	Results	Individual Judgment																																
1	General requirements	—	<p>Materials:</p> <p>a) Materials in direct contact with the face should be harmless to the skin.</p> <p>(b) The filter media shall be harmless to humans.</p> <p>(c) The material used shall have sufficient strength to withstand breakage or deformation during its normal service life.</p> <p>Structure:</p> <p>(a) should not be susceptible to structural damage and the design, composition and installation of the components should not pose any hazard to the user.</p> <p>(b) The headband should be designed to be adjustable and easy to wear and remove, should securely fasten the mask to the face and should be worn without visible compression or pain, and the headband design of the replaceable half mask and full mask should be replaceable.</p> <p>(c) Should have as small a dead space and a large field of view as possible.</p> <p>(d) The disposable mask shall be constructed to ensure a close fit to the face and shall be free from deformation during its service life.</p>	<p>Structure:</p> <p>a) That is not susceptible to structural damage and the design, composition and installation of the components do not pose any hazard to the user.</p> <p>b) The headband is designed to be adjustable for easy wearing and removal, to hold the mask securely to the face, and to be worn without significant compression or pressure pain.</p> <p>c) Has a smaller dead space and a larger field of view.</p>	—																																
2	External Inspection	—	The surface shall be free from breakage, deformation and other obvious defects, the material and construction of the components shall be able to withstand normal conditions of use and the temperature, humidity and mechanical shocks that may be encountered, the headband shall be adjustable and the headband design of the replaceable mask shall be replaceable. Parts should not be dislodged, damaged or deformed after pre-treatment with temperature and humidity.	Meet the requirements.	Qualified																																
3	Respiratory resistance	Pa	Total inhalation resistance should not exceed 350 and total exhalation resistance should not exceed 250 for each sample.	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2" style="text-align: center;">Sample No.</th> <th colspan="2" style="text-align: center;">MV</th> </tr> </thead> <tbody> <tr> <td rowspan="2" style="text-align: center;">Total inspiratory resistance</td> <td style="text-align: center;">Unprocessed</td> <td style="text-align: center;">21<sup>#</sup></td> <td style="text-align: center;">122.4</td> </tr> <tr> <td style="text-align: center;">Pretreatment</td> <td style="text-align: center;">22<sup>#</sup></td> <td style="text-align: center;">119.2</td> </tr> <tr> <td rowspan="2" style="text-align: center;">Total exhalation resistance</td> <td style="text-align: center;">Unprocessed</td> <td style="text-align: center;">25<sup>#</sup></td> <td style="text-align: center;">119.4</td> </tr> <tr> <td style="text-align: center;">Pretreatment</td> <td style="text-align: center;">26<sup>#</sup></td> <td style="text-align: center;">117.6</td> </tr> <tr> <td rowspan="2" style="text-align: center;">Total inspiratory resistance</td> <td style="text-align: center;">Unprocessed</td> <td style="text-align: center;">23<sup>#</sup></td> <td style="text-align: center;">107.4</td> </tr> <tr> <td style="text-align: center;">Pretreatment</td> <td style="text-align: center;">24<sup>#</sup></td> <td style="text-align: center;">99.5</td> </tr> <tr> <td rowspan="2" style="text-align: center;">Total exhalation resistance</td> <td style="text-align: center;">Unprocessed</td> <td style="text-align: center;">27<sup>#</sup></td> <td style="text-align: center;">92.4</td> </tr> <tr> <td style="text-align: center;">Pretreatment</td> <td style="text-align: center;">28<sup>#</sup></td> <td style="text-align: center;">91.9</td> </tr> </tbody> </table>	Sample No.		MV		Total inspiratory resistance	Unprocessed	21 <sup>#</sup>	122.4	Pretreatment	22 <sup>#</sup>	119.2	Total exhalation resistance	Unprocessed	25 <sup>#</sup>	119.4	Pretreatment	26 <sup>#</sup>	117.6	Total inspiratory resistance	Unprocessed	23 <sup>#</sup>	107.4	Pretreatment	24 <sup>#</sup>	99.5	Total exhalation resistance	Unprocessed	27 <sup>#</sup>	92.4	Pretreatment	28 <sup>#</sup>	91.9	Qualified
Sample No.		MV																																			
Total inspiratory resistance	Unprocessed	21 <sup>#</sup>	122.4																																		
	Pretreatment	22 <sup>#</sup>	119.2																																		
Total exhalation resistance	Unprocessed	25 <sup>#</sup>	119.4																																		
	Pretreatment	26 <sup>#</sup>	117.6																																		
Total inspiratory resistance	Unprocessed	23 <sup>#</sup>	107.4																																		
	Pretreatment	24 <sup>#</sup>	99.5																																		
Total exhalation resistance	Unprocessed	27 <sup>#</sup>	92.4																																		
	Pretreatment	28 <sup>#</sup>	91.9																																		
4	Vision	—	Lower field of view $\geq 60^\circ$	$64^\circ$	Qualified																																
5	Dead Space	—	Volume percentage of carbon dioxide shall be $\leq 1\%$	Volume percentage of carbon dioxide is 0.6%.	Qualified																																



## Testing Results

STFWT20204824

Page 3 of 5

Serial	Test Items	Unit	Requirement	Results	Individual Judgment		
6	Filtration efficiency/% (NaCl particulate matter)	---	KN95: $\geq 95.0$	Sample No.	MV	Qualified	
				1 <sup>#</sup>	Initial		97.7
					Loaded		97.5
				2 <sup>#</sup>	Initial		97.9
					Loaded		97.7
				3 <sup>#</sup>	Initial		98.0
					Loaded		97.7
				4 <sup>#</sup>	Initial		98.2
					Loaded		98.0
				5 <sup>#</sup>	Initial		97.9
					Loaded		97.7
				6 <sup>#</sup>	Initial		98.4
					Loaded		98.2
				7 <sup>#</sup>	Initial		98.2
					Loaded		98.0
				8 <sup>#</sup>	Initial		98.4
					Loaded		98.1
				9 <sup>#</sup>	Initial		98.6
					Loaded		98.4
				10 <sup>#</sup>	Initial		98.4
Loaded	98.1						
11 <sup>#</sup>	Initial	97.6					
	Loaded	97.4					
12 <sup>#</sup>	Initial	97.8					
	Loaded	97.6					
13 <sup>#</sup>	Initial	97.6					
	Loaded	97.4					
14 <sup>#</sup>	Initial	97.9					
	Loaded	97.7					
15 <sup>#</sup>	Initial	98.0					
	Loaded	97.8					
7	Headband	---	Each headband, buckle and other adjusting parts of the mask should not slip or break when subjected to a 10N tension for 10s.	Not processed: 31#Each headband, buckle and other adjusting parts of the 1# mask did not slip or break after being subjected to 10N of tension for 10s. After pretreatment: Each headband, buckle and other adjusting parts of the 5# mask did not slip or break after being subjected to 10N tension for 10s.	Qualified		

151

## Testing Results

STFWT20204824

Page 4 of 5

Serial	Test Items	Unit	Requirement	Results	Individual Judgment			
8	Breath valve cover	---	The exhalation valve cover is subjected to 10N axial tension for 10s, there should be no slippage, fracture and deformation.	Without this component, this item will not be evaluated	---			
9	Breathing valve air tightness	---	Each sample shall be free from one of the following conditions.(a) When the pumping flow rate has reached 500 mL/min, the system negative pressure is less than 1180 Pa.(b) The time for the expiratory valve to return to atmospheric pressure is less than 20s.	Without this component, this item will not be evaluated	---			
10	Leakage/% (Total Inward Leakage of Disposable Mask)	---	Filter media level	Based on Til of each action (i. e. 10 persons and 5 actions) , TIL of at least 46 actions in 50 actions	Unproc essed	16 <sup>#</sup>	There are 47 actions with a TIL less than 11.	Qualified
						17 <sup>#</sup>	There are 47actions with a TIL less than 11.	
						18 <sup>#</sup>	There are 47actions with a TIL less than 11.	
						19 <sup>#</sup>	There are 47actions with a TIL less than 11.	
						20 <sup>#</sup>	There are 47actions with a TIL less than 11.	
			KN95	< 11	Pretreat ment	21 <sup>#</sup>	There are 47 actions with a TIL less than 11.	
						22 <sup>#</sup>	There are 47 actions with a TIL less than 11.	
						23 <sup>#</sup>	There are 47actions with a TIL less than 11.	
						24 <sup>#</sup>	There are 47 actions with a TIL less than 11.	
						25 <sup>#</sup>	There are 47 actions with a TIL less than 11.	
			Filter media level	Total Inward Leakage for at least 8 out of 10 subjects when evaluated on a total person basis	Unproc essed	16 <sup>#</sup>	There were 9 people with a TIL less than 8.	
						17 <sup>#</sup>	There were 9 people with a TIL less than 8.	
						18 <sup>#</sup>	There were 9 people with a TIL less than 8.	
						19 <sup>#</sup>	There were 9 people with a TIL less than 8.	
						20 <sup>#</sup>	There were 9 people with a TIL less than 8.	
KN95	< 8	Pretreat ment	21 <sup>#</sup>	There were 9 people with a TIL less than 8.				
			22 <sup>#</sup>	There were 9 people with a TIL less than 8.				
			23 <sup>#</sup>	There were 9 people with a TIL less than 8.				
			24 <sup>#</sup>	There were 9 people with a TIL less than 8.				
			25 <sup>#</sup>	There were 9 people with a TIL less than 8.				

黃

## Testing Results

STFWT20204824

Page 5 of 5

Serial	Test Items	Unit	Requirement	Results	Individual Judgment
11	Flammability	—	The parts exposed to the flame should not burn after being removed from the flame; if they do, the renewal time should not exceed 5s.	Not processed: 1 <sup>#</sup> , 2 <sup>#</sup> The parts of the specimen exposed to the flame did not burn after being removed from the flame. After pretreatment: 3 <sup>#</sup> , 4 <sup>#</sup> The parts of the specimen exposed to the flame did not burn after being removed from the flame.	Qualified
12	Connect and connect parts	—	Replaceable type filter element and masks, breathing tube and filter element and mask all connections and links between parts, in 10 n under axial tension and lasts 10 s, should not appear slippage, fracture or deformation.	With the abandoned mask, do not check this	—

### Test sample picture



The following blank